

## ACQUITY UPLC H-Class PLUS System with 2D Technology

The Waters™ ACQUITY™ UPLC™ H-Class PLUS System with 2D Technology allows chemists to increase sensitivity and selectivity, eliminate unwanted interferences, characterize the most complex samples, and perform separations that are normally incompatible with a mass spectrometer by adding a second reversed-phase separation to the experiment. The system is comprised of one Quaternary Solvent Manager (Injection Pump), one Binary Solvent Manager (Analytical Pump), a Sample Manager with Flow-Through Needle (SM-FTN-H), and a Column Manager.

### BINARY SOLVENT MANAGER (BSM)

|                              |   |
|------------------------------|---|
| Number of solvents           | Up to four, in any combination of two: A1 or A2 and B1 or B2  |
| Solvent conditioning         | Vacuum degassing: one channel per solvent, and one channel for Sample Manager wash solvent  |
| Settable flow rate range     | 0.001 to 2.000 mL/min, in 0.001 mL increments   |
| Compressibility compensation | Automatic and continuous  |
| Plunger seal wash            | Integrated, active, programmable  |
| Gradient profiles            | 11 gradient curves (including linear, step [2], concave [4], and convex [4])  |
| Wet prime                    | Automatic   |
| Maximum operating pressure   | 15,000 psi up to 1 mL/min, 9000 psi up to 2 mL/min per pump, not more than 15,000 psi total   |
| Composition accuracy         | ±0.5% absolute from 5% to 95%, 0.5 to 2.0 mL/min  |
| Composition precision        | 0.15% RSD or ±0.04 min SD, whichever is greater (from 0.2 to 2.0 mL/min)  |
| Flow precision               | 0.075% RSD or ±0.02 min SD, or 1.00 s for run times less than 1.00 min based on retention time or volumetric measures (0.50 to 2.00 mL/min) |
| Primary wetted materials     | 316 stainless steel, UHMWPE, sapphire, ruby, fluoropolymer, DLC, PEEK and PEEK blend, titanium alloys                                       |

## QUATERNARY SOLVENT MANAGER (QSM)

|                                      |  |
|--------------------------------------|--|
| Number of solvents                   | Blend up to four solvents in any combination (standard)<br>Expanded solvent choices with optional six-port solvent select valve  |
| Maximum operating pressure           | 15,000 psi up to 1.0 mL/min, 9000 psi up to 2.0 mL/min (firmware version 1.5x and earlier)<br>15,000 psi up to 1.0 mL/min, 7800 psi up to 2.2 mL/min (firmware version 1.6x and later)   |
| Solvent degassing                    | Integrated vacuum degassing, four chambers<br>One additional chamber for the SM-FTN-H purge solvent  |
| Solvent blending                     | Automated, on-line pH, ionic strength, and organic modifier blending from pure solvents with Auto-Blend Plus™ Technology   |
| Gradient formation                   | Low-pressure mixing, quaternary gradient   |
| Gradient profiles                    | 11 gradient curves [including linear, step (2), concave (4), and convex (4)]   |
| Settable flow rate range             | 0.010 to 2.000 mL/min, in 0.001 mL increments (firmware version 1.5x and earlier)<br>0.010 to 2.200 mL/min, in 0.001 mL increments (firmware version 1.60)<br>0.001 to 2.200 mL/min in 0.001 mL increments (firmware version 1.65 and later) |
| Primary check valve                  | Intelligent Intake Valve ( <i>i</i> <sup>2</sup> Valve), standard<br>Passive check valve (optional)  |
| Pressure pulsation†                  | ≤1.0% or 25 psi, whichever is greater  |
| Flow accuracy†                       | ±1.0% at 0.5 to 2.0 mL/min using 100% A (with <i>i</i> <sup>2</sup> Valve)   |
| Flow precision†                      | ≤0.075% RSD or ±0.01 min SD, whichever is greater, based on six replicates (with <i>i</i> <sup>2</sup> Valve)  |
| Composition ripple† (baseline noise) | ≤1.0 mAu (≤0.1 mAU with optional 250 µL mixer) (with <i>i</i> <sup>2</sup> Valve)  |
| Composition accuracy†                | ±0.5% absolute (full scale) from 5% to 90% from 0.5 to 2.0 mL/min (with <i>i</i> <sup>2</sup> Valve)   |
| Composition precision†               | ≤0.15% RSD or ±0.02 min SD, whichever is greater, based on six replicate injections ( <i>i</i> <sup>2</sup> Valve)   |
| Compressibility compensation         | Automatic and continuous   |
| Priming                              | Wet priming can run at flow rates up to 4 mL/min   |
| Pump seal wash                       | Equipped with a wash system to flush the rear of the high pressure seal and the plunger  |
| Flow ramping                         | Range: 0.01 to 30.00 min to reach 2.0 mL/min<br>Default: 0.45 min to reach 2.0 mL/min  |
| Primary wetted materials             | 316 stainless steel, PPS, fluoropolymer, fluoroelastomer, UHMWPE blend, sapphire, ruby, zirconia, Nitronic 60, DLC, PEEK and PEEK blend, titanium alloy  |

## SAMPLE MANAGER-FTN (SM-FTN-H)

|                                      |   |
|--------------------------------------|---|
| Injection volume range               | 0.1 to 10.0 $\mu\text{L}$ as standard<br>Up to 1000.0 $\mu\text{L}$ with optional extension loops   |
| Accuracy (aspiration)                | $\pm 0.2$ $\mu\text{L}$ , measured by fluid weight removed from vial with 10.0 $\mu\text{L}$ injections averaged over 20 injections using standard 100 $\mu\text{L}$ syringe  |
| Linearity <sup>†</sup>               | $\geq 0.999$ (standard needle)  |
| Precision <sup>†</sup>               | $\leq 0.25\%$ RSD, 5 to 100 $\mu\text{L}$   |
| Number of sample plates              | Any two of the following: <ul style="list-style-type: none"> <li>▪ 96 and 384 microtiter plates</li> <li>▪ 48 position 2.00-mL vial plates</li> <li>▪ 48 position 0.65-mL micro-centrifuge tube plates</li> <li>▪ 24 position 1.50-mL micro-centrifuge tube plates</li> </ul> |
| Maximum sample capacity              | 768 in two 384-well plates, or 96 in 2-mL vial holders.<br>Additional positions for dilution functions  |
| Sample compartment                   | 4.0 to 40.0 $^{\circ}\text{C}$ , settable in 0.1 $^{\circ}\text{C}$ increments  |
| Temperature accuracy                 | $\pm 0.5$ $^{\circ}\text{C}$ at sensor  |
| Temperature stability                | $\pm 1.0$ $^{\circ}\text{C}$ at sensor  |
| Sample manager heat time             | $\leq 30$ min ambient-40 $^{\circ}\text{C}$   |
| Sample manager cool time             | $\leq 60$ min ambient-4 $^{\circ}\text{C}$  |
| Injection needle wash                | Integrated, active, programmable  |
| Minimum sample required              | 3 $\mu\text{L}$ residual, using Waters total recovery 2-mL vials (zero offset)  |
| Sample carryover <sup>†</sup>        | $\leq 0.002\%$ caffeine (UV)<br>$\leq 0.002\%$ sulphadimethoxine (MS)   |
| Advanced Sample Manager capabilities | Auto-dilution, auto-addition, and load-ahead  |
| Primary wetted materials             | 316 stainless steel, gold plated stainless steel, polyimide, PEEK blend, DLC  |

## COLUMN MANAGEMENT (CM-A)

|   |   |
|---|---|
| Column capacity                             | CM-A: Two columns, as standard (maximum length of 150 mm with filter or guard column), or four columns (maximum length of 50 mm) can be supported with optional tubing kit, up to 4.6 mm internal diameter (I.D.) |
| Multidimensional valves                     | Two six-port, two-position valves (CM-A only)   |
| Column compartment(s) temperature range     | 4.0 to 90.0 °C, settable in 0.1 °C increments<br>Two independent heat/cool zones per module   |
| Column compartment(s) temperature accuracy  | ±0.5 °C   |
| Column compartment(s) temperature stability | ±0.3 °C   |
| Column compartment heat time                | ≤15 min ambient-60 °C   |
| Column compartment cool time                | ≤15 min 60–20 °C  |
| Solvent conditioning                        | Active pre-heating as standard  |
| Column tracking                             | eCord™ Technology column information management tracks and archives column usage history for one column   |

## INSTRUMENTAL CONTROL

|                        |   |
|------------------------|---|
| External communication | Ethernet interfacing via RJ45 connection to host PC with BSM, or Column Manager and UPLC detectors and mass spectrometers                             |
| Event inputs/outputs   | Rear panel contact closure and/or TTL inputs/outputs  |
| External control       | MassLynx® version 4.1 with OpenLynx™ Open Access, with specific SCN releases  |
| User diagnostics       | Available through software on host PC; system control via console software  |
| Unattended operation   | Leak sensors on supported modules, full diagnostic data captured through console software   |
| Connections INSIGHT™   | Provides real-time monitoring and automatic notification of instrument performance and diagnostic information allowing for quicker problem resolution |

## ENVIRONMENTAL

|                             |                                   |
|-----------------------------|-----------------------------------|
| Acoustic noise              | <65 Db                            |
| Operating temperature range | 4.0 to 40.0 °C (39.2 to 104.0 °F) |
| Operating humidity range    | 20% to 50%, non-condensing        |

## POWER REQUIREMENTS

|         |                |
|---------|----------------|
| Voltage | 100 to 240 VAC |
|---------|----------------|

|           |             |
|-----------|-------------|
| Frequency | 50 to 60 Hz |
|-----------|-------------|

## PHYSICAL DIMENSIONS

|                                  |                         |
|----------------------------------|-------------------------|
| ACQUITY UPLC H-Class PLUS System | Width: 83.8 cm (33 in.) |
|----------------------------------|-------------------------|

|                     |                             |
|---------------------|-----------------------------|
| with 2D Technology: | Height: 103.4 cm (40.7 in.) |
|---------------------|-----------------------------|

|                                    |                         |
|------------------------------------|-------------------------|
| ACQUITY UPLC Sample Manager-FTN-H, | Depth: 86.4 cm (34 in.) |
|------------------------------------|-------------------------|

Binary Solvent Manager, Quaternary

Solvent Manager, and Column Manager

Note: dimensions are listed with only components listed above

## ORDERING INFORMATION

## PART NUMBER

|   |
|---|
| ACQUITY UPLC H-Class PLUS System with 2D Technology |
|---|

|           |
|-----------|
| 176015132 |
|-----------|

*\* For specific test conditions, contact your Waters sales representative.*

# Waters

**THE SCIENCE OF WHAT'S POSSIBLE.™**

Waters, The Science of What's Possible, ACQUITY, Auto-Blend Plus, Connections INSIGHT, eCord, MassLynx, OpenLynx, and UPLC are trademarks of Waters Corporation. All other trademarks are the property of their respective owners.

©2018 Waters Corporation. Produced in the U.S.A. April 2018 720006263EN LM-PDF

Waters Corporation

34 Maple Street

Milford, MA 01757 U.S.A.

T: 1 508 478 2000

F: 1 508 872 1990

[www.waters.com](http://www.waters.com)